


StickyScan™ 100



LCD monitor, scanner, keyboard and mouse

**Production Test Method -
Quick test with little training
required.**

Description

The StickyScan100 is specifically designed to be used in conjunction with the Pulmac MasterScreen to generate a stickies count based upon screening contaminants, displaying them on a 22 cm in diameter filter paper, preparing the display, then scanning. Depending on pulp and application a 0.004", 0.006" or 0.008" slotted screen plate may be used.

Benefits

The StickyScan100 will generate simple and effective dirt data for feedback to production operators for abatement efforts. The StickyScan100 is easy to set up, easy to train and simple to interpret data.



Pulmac Systems International Inc.
35 Dorset Lane, Suite 201
Williston, Vermont
05495 USA

Phone: + 802.223.2663
Fax: + 802.223.2664
Email: info@pulmac.com
URL: www.pulmac.com

StickyScan™ 100

Fast & Easy

Full scan and data output takes less than 1 minute. Just click on "Test" button.

Count	Total Area	Mean Area	PPM
163	30.3	0.188	1062

Location	Description	Sample Date/Time	Grams	Count	Total Area	Mean Area	PPM
Recycled Tissue	PM10 Stuff Box Late in System	11/12/2008 11:00:51	1000	163	30.3	0.188	1062
Recycled Tissue	PM10 Stuff Box Late in System	11/10/2008 18:50:08	1000	304	53.5	0.178	1877
Recycled Tissue	PM10 Stuff Box Late in System	11/10/2008 15:55:10	1000	230	35.0	0.150	1350
Recycled Tissue	FD1 Slush Tank Early in System	11/10/2008 15:04:12	1000	480	82.1	0.171	2670
Recycled Tissue	FD1 Slush Tank Early in System	11/10/2008 15:00:11	1000	558	87.9	0.151	3684
Recycled Tissue	PM10 Stuff Box Late in System	11/10/2008 18:21:09	1000	284	48.8	0.189	1624
Recycled Tissue	PM10 Stuff Box Late in System	11/10/2008 13:12:20	1000	200	29.9	0.149	1047
Recycled Tissue	PM10 Stuff Box Late in System	11/09/2008 17:55:21	1000	1050	23.2	0.022	515
Recycled Tissue	PM10 Stuff Box Late in System	11/07/2008 14:29:16	1000	163	14.4	0.090	564
Recycled Tissue	FD1 Slush Tank Early in System	11/07/2008 14:22:11	1000	230	15.2	0.066	532
Recycled Tissue	PM10 Stuff Box Late in System	11/07/2008 12:54:29	1000	238	17.0	0.071	597

StickyScan100 screenshot of monitor before running a test.

StickyScan100

Specifically designed for use with Pulmac MasterScreen.

System Features:

INPUT

- Enter location: Drop down box allows location to be selected from a list or entered with keyboard.
- Enter Description: Drop down box allows selection of description from a list or entered with keyboard.
- Enter Time/Date: Enter with keyboard or use Default of current time/date.
- Enter sample size: Enter with keyboard or use default of 1000 g (data is normalized to 1 kg).

OUTPUT

- Current test data of "Count", "Total Area", "Mean Area", "PPM" is prominently displayed.
- Image of current scan is displayed.
- Last 11 test results are displayed. Entire history is available using navigation buttons.
- Summary and distribution data are automatically copied to supplied USB Flash Drive.

CALIBRATION

- Verify calibration using supplied reference.

PACKAGE

- PC Box, LCD Monitor, Scanner, USB Flash Drive
- Preinstalled operating and application software
- Back up CD and reference sheet

Represented By:

Pulmac Systems International Inc.
35 Dorset Lane, Suite 201
Williston, Vermont
05495 USA

Phone: + 802.223.2663

Fax: + 802.223.2664

Email: info@pulmac.com

URL: www.pulmac.com