

MasterScreen™



XLQ MasterScreen™

Rapid Results -

Displays in less than 10 minutes

Description

The MasterScreen is a low consistency screening device designed to mechanically separate fibers from contaminants, including Stickies, Dirt, Shives, vessel cell elements, plastics and hot melts. Contaminants are deposited automatically onto a filter paper for display and analysis.

Benefits

The MasterScreen provides a rapid and repeatable method for the collection of both Shives and/or Stickies. For Shives, it allows for the pinpointing of problems that occur in areas such as digesters, cleaners, refiners, grinders and screens. For Stickies, it allows for the pinpointing of problems that occur in repulpers, cleaners and screening. This data will support reducing the frequency of Dirt, Shives or Stickies in the final product, leading to increased profitability.

Bottom-line: Increased profitability through troubleshooting, optimization and monitoring.



PulmacSystemsInternational Inc.
35 Dorset Lane, Suite 201
Williston, Vermont
05495 USA

Phone: + 802.223.2663
Fax: + 802.223.2664
Email: info@pulmac.com
URL: www.pulmac.com

MasterScreen™



System Features:

- Pressurized screening chamber adjustable to 15 psi
- Screens up to 100 g (oven dried equivalent)
- Results between 10 - 15 minutes
- Laser cut screen plates available in two standard sizes (.004" and .006") and custom between .003" - .014"
- AutoFilter for quick deposit of contaminants
- AutoFilter can be customized to accommodate various sized filter paper
- AutoFilter cartridge is easily repaired with "bolt in" screen
- Lid on feed tank localizes water spray
- Large feed tank to accept samples up to 100 grams, oven dried equivalent
- Feed tank air mix to ensure delivery of uniform consistency
- XLQ valve (option) between screening chamber and feed tank
- Multi program feature including MS A, XLQ, SA, and SCR N
- Adjustable feed tank level control.

Related Test Methods:

- PAPTAC C.12P Sep 03 Laboratory Screening of Pulp (Pulmac-Type Instrument)
- TAPPI T 274 sp04 Laboratory Screening of Pulp (MasterScreen - Type Instrument)
- INGEDE Method 4 99-12 Analysis of Macro Stickies in Deinked Pulp (DIP)
- ASTM D6148-97 (2002) Standard Practice for the Separation and Examination of Stickies
- ISO 15360 - 2:2001 Recycled Pulps - Estimation of Stickies and Plastics image analysis method
- TAPPI T 277 pm-99 Macro Stickies Content in Pulp: The "Pick-Up" Method.

XLQ Valve Option



Auto Filter

Represented By:

PulmacSystemsInternational Inc.
35 Dorset Lane, Suite 201
Williston, Vermont
05495 USA

Phone: + 802.223.2663

Fax: + 802.223.2664

Email: info@pulmac.com

URL: www.pulmac.com

Shipping Specs		Height	Width	Depth
Crated Weight & Size	150 kg	84 cm	219 cm	89 cm

Service and Hardware Requirements:

- 40 L per minute of potable filtered water
- 3/4" NPT male fitting for water supply
- 1 1/2" NPT male fitting for drain
- 240 V, 50 Hz electrical service
- Compressed air at 60 l/m at 5 bar
- 10 mm slip fitting for air supply